International Academic Institute for Science and Technology

International Academic Journal of Science and Engineering
Vol. 12, No. 1, 2025, pp. 16-20.

ISSN 2454-3896

International Academic Journal of Science and Engineering

www.iaiest.com

Investigating the Role of Fractals in Enhancing MRI Image Compression Techniques and Signal Processing Efficiency

Veer Saxena [a], Preeti Mhatre [b]

[a] *Cognitive Architectures Unit, Jhansi School of Artificial Reasoning, India.*

[b] *Cognitive Architectures Unit, Jhansi School of Artificial Reasoning, India.*

Abstract

Magnetic Resonance Imaging (MRI) provides significant medical benefits by aiding physicians in clinical staging and predicting surgical extent. A substantial quantity of MRI images necessitates considerable storage capacity and transmission rates within the system for offline preservation and diagnosis from afar. The enlargement of superior-quality MRI pictures is very research-focused. Current MRI image compression algorithms that achieve substantial compression ratios result in loss of data regarding tumors, which can lead to misinterpretation; conversely, systems with low compression ratios fail to produce the desired outcomes. This research proposes a rapid fractal-based reduction technique for MRI pictures. Initially, three-dimensional (3D) MRI pictures are transformed into a two-dimensional (2D) image series, enabling the sequence to utilize fractal compression. Range and area blocks are categorized based on the intrinsic spatiotemporal resemblance of three-dimensional objects. Applying self-similarity decreases the number of blocks in the comparing pool, enhancing the comparing speed of the suggested approach. A residual correction approach is implemented to attain high-quality decompression of MRI images using compressing. The experiments indicate that the reduction speed has increased by 2 to 3 times, and the Peak Signal Noise Ratio (PSNR) has increased by about 10. The suggested approach resolves the conflict between significant compression coefficients and the standard of MRI medical pictures.

Keywords: Fractals, Magnetic Resonance Imaging, Image Compression, Signal Processing.

DOI: 10.71086/IAJSE/V12I1/IAJSE1204
Received: 16 December 2024/Revised: 20 January 2025/Accepted: 15 February 2025/Published: 31 March 2025

Introduction

Magnetic Resonance Imaging (MRI) (Reda et al., 2021) has emerged as a more dynamic and swiftly advancing medical research and diagnostics domain. Medical pictures illustrate the interior architecture of the human body comprehensibly, offering practitioners precise and fundamental insights into anatomy, disease, and function. Standard methods of imaging comprise MRI, computed tomography (CT) (Withers et al., 2021), ultrasound (US), computed radiography (CR) (Nagesha et al., 2022), and digital subtraction angiography (DSA) (Posa et al., 2022). As medical imaging advances, particularly in device resolution, the volume of medical image information will persist and increase.

Medical photo compression techniques are often categorized into lossless and lossy compression types (Elakkiya & Thivya, 2022). Lossless compression delivers healthcare diagnostics with image data of identical quality to the original image. The compression percentage of lossless compression is typically poor, making it challenging to satisfy the actual transmission standards for medical pictures. Lost compression achieves a more excellent ratio of compression by discarding certain information. An increase in compressing necessarily results in deterioration in medical photographs.

MRI medical pictures exhibit substantial temporal and spatial connections. Three analogous image blocks, labeled 1, 2, and 3, are separately obtained from three contiguous MRI slices. The spatiotemporal correlations indicate that in an MRI slice, the neighboring fields of a specific pixel (region) within a layer exhibit significant similarity to that pixel (area) over numerous preceding and subsequent layers. This spatiotemporal relationship encapsulates substantial (local) self-similar data inside MRI pictures. Compressing this category of medical images is deemed appropriate to utilize the concept of fractal reduction (Zerva et al 2023).

Fractal picture compression relies on the Repeated Function System, discarding the image content while preserving just the self-similarity variables of the local image features to achieve data reduction. It offers the benefits of a strong compression ratio, scalable rebuilding, and rapid decoding. Conventional fractal compression methods inadequately account for the temporal relationships inherent in MRI scans, leading to suboptimal and unsatisfactory outcomes. This research proposes a rapid fractal-based reduction technique for MRI pictures.

Background

Fractal compression of pictures was initially introduced in a study. Numerous further papers have emerged on this subject. This paper outlines the design of geometric compression of images. Some articles achieve a high compression ratio, while others demonstrate reduced encoding time. The research posits that these folders can assist forensic scientists in supplying proof for files accessed from the WinRAR archives. This evidence will be invaluable if the person denies accessing the files. WinRAR software provides encryption and reduction.

The study presented an approach that aggregates related weights of specific characteristics in both the spatial and descriptive domains, facilitating the retrieval of images from an archive containing many photos (Ma et al 2021). The efficacy of specific local characteristics can be enhanced with minimal computational cost. Experimental findings from evaluations indicate that this strategy exhibits outstanding efficiency with reduced computational requirements and enhanced retrieval accuracy. The study developed an image reduction method based on regionally varying image characteristics. Based on this knowledge, a picture is segmented into blocks, and polynomial approximations are employed to deconstruct the image into less compressed information. The transfer rate is enhanced via Huffman encoding.

Studies indicate that the computational difficulty of fractal compression of images primarily arises from the extensive comparisons required to identify matching domains related to block spans within the picture (Song et al 2024). This article examines the multi-faceted computing efficiency of branch research about domain-related data retention. Cabin subdomains will be organized in a B++ tree and arranged according to one or more chosen local characteristics of each site.

Enhanced MRI Image Compression using Fractals

The algorithmic flow for implementing the suggested compression strategy is illustrated in Fig. 1. The initial phase of the proposed approach involves delineating the Region of Interest (ROI) and Non-ROI (NROI) regions by morphological dissection. The morphological separation converts an input brain picture into a picture in grayscale. Building components are allocated. The supplied image is processed with the designated structural element utilizing

morphological operations such as Destruction and Dilation. In the generated binary image, pixels with a value of 1 are substituted with their corresponding native pixel values from the input picture, resulting in the ROI part of the picture. To acquire the NROI part, the binary picture is turned upside down. The pixels with an integer of 0 are substituted with their corresponding native pixel values from the input picture.




Fig. 1: Workflow of the Image Compression

After segmentation, the ROI region is subjected to compression with no loss via the Burrows-Wheeler compression technique. The primary concern is its computational difficulty and the duration required for rearrangement. The ROI picture can be diminished by segmenting it into smaller blocks and applying arrangements. The Huffman encoding algorithm has been utilized for final encoding. The NROI area has been reduced with a mixed fractal approach. The primary concern with conventional blocks is the extended encoding duration. An innovative fractal encoding technique, which combines fractal coding with other coding techniques, is employed to address this. Quadtree reduction and zero mean brightness level with Huffman encoding compress the NROI area. The original image is initially segmented into range blocks using the Quad-tree partitioning technique, which splits an image square into four equal-sized rectangular blocks. Each block is evaluated to determine if it satisfies specific homogeneity requirements. Should the block satisfy the conditions, it remains undivided. If it fails to match the criteria, it is broken into four data blocks, and the test requirements are assigned to those pieces.

A fractured compression system designed for MRI scans leverages the self-similarity characteristic of MRI pictures and consists of three stages: Education, compressing, and Decoding. The primary objective of this decompression method is to minimize the search area and expedite the encoding duration. The subsequent sub-sections elaborate on these three stages.

Phase A: Instruction

During this step, analogous MRI scans are utilized as input, independently segmenting each image into range blocks and domain chunks. In the preliminary phase, the range block of a single MRI scan is contrasted with the area bricks of the same picture to identify the most suitable domain blocks for the specific range block, resulting in graphic coding. The expertise system has been trained using the range indexes and their optimally matched area chunks. The duration of the training procedure will not be included in the encoded time.

Phase B: Encryption

In the encoding step, the expert systems sequentially input the index of each range region of the MRI picture and generate a corresponding set of area blocks for the corresponding ranging blocks. The lookup will be conducted solely within the resultant collection of domain names. The search space is decreased, resulting in an accelerated encoding time.

Phase C: Decoding Phase

The reverse decoding method involves applying modification settings to a starting image with a mean worth, which converges to the fractal picture in fewer than 10 rounds. The input MRI scan is first divided into fixed-size segments

known as Range Stops. A block processing operation diminishes the newly created image by averaging the brightness of four adjacent pixels. This results in three vectors of the recording stage for consideration. A solitary block is initially selected from the diminished picture. The associated best-matched subdomain block gauge, scaling variable, and offset variables are extracted from the matrices. The values of the pixels of the diminished area block are positioned inside the range determined by the orientation data following scaling and offsets. All of the possible blocks comprise a singular iteration. The decompressed images will be generated after one to 10 rotations.

Results

The suggested ROI-based reduction technique has been executed and evaluated on medical pictures, including ultrasound, CT, MRI, and X-ray images from the Medpix collection. The data illustrates the compression efficacy of the suggested approach across various medical photos. The results indicate variations in ultrasound pictures of differing sizes. A mean PSNR of 37.2 dB is achieved with a compression factor of 4.6. A median space-saving rate of 78.5 is achieved, which is crucial for distribution and storage.

The research compresses the initial four layers of the three separate MRI scans. A combinatorial technique is employed that integrates the four levels of pictures and reduces them. The categorizing of area bricks and range bricks is conducted worldwide. The "codebook" has been augmented from a single-layer picture to a four-layer image, enhancing matching precision. Initially, the research reduce the 4-layer picture without employing a residual compensating system, followed by compression utilizing the enhanced fractal coding approach with a residual compensating system. The domain bricks and range blocks are categorized into ten equivalent classes upon amalgamation. Table 1 presents the effectiveness of the three approaches, with the Peak Signal Noise Ratio (PSNR) being the mean of the 4-layer picture.

Table 1: Performance Analysis

Method		Traditional	Without RCM	With RCM
Screening	T	121.3	46.32	48.24
	PSNR	28.43	27.32	37.42
Complete 1 Year	T	127.32	43.25	49.21
	PSNR	29.43	28.54	38.65
Complete 2 Years	T	120.43	42.65	51.26
	PSNR	30.14	30.62	40.32

Table 1 demonstrates that the picture quality achieved by the suggested residual compensation methods is ideal. Despite having a longer compressing time than those indicated and lacking an extra compensation system, it remains significantly less than the time needed by conventional approaches, and its PSNR is the highest of the four. The duration for processing four layers of MRI scans merged and compressed is quadruple that of one layer image. All range and area blocks are categorized into ten categories, thus maintaining a speed-up effect. The reduction duration is 2.5 times longer than that of the conventional approach. The implementation markedly enhances the quality of reconstructed pictures. It should adequately rectify the mistake loss.

Conclusion

The swift advancement of medical imaging equipment has resulted in the generation of substantial three-dimensional health information, including MRI, CT, and three-dimensional ultrasonography. The considerable amount of three-dimensional imaging data incurs significant transportation and storage costs during diagnoses and therapy. Effective compression of medical images decreases the volume of data transmitted. The compressed information is stored and sent. It conserves storage capacity, enhances transmission speed, and reduces transmission duration during distant diagnostics. Fostering the advancement of healthcare.

- The research suggests a rapid MRI compression technique utilizing fractal principles to address the abovementioned issues. The dimensional MRI image is initially transformed into a two-dimensional sequential image. The series employs an image-based fractal reduction technique for reduction. Secondly, range bricks and area blocks are categorized based on spatiotemporal resemblance characteristics. A range block requires searching solely for the ideal matching block inside a specific category of domain blocks. Expedite by decreasing the size of the matching group. The proposed procedure's validity is confirmed.
- In the future, the research must enhance operations as outlined below.
- Consider recruiting additional volunteers for future investigations and doing statistical tests.

- Develop additional elements to improve the delineation of domain division categorization.
- Suggest a more efficient medical photograph reduction algorithm utilizing cloud computing, machine learning, and other relevant concepts.
- Initiate improvements to the compression method to enhance its error tolerance.

References

Reda, R., Zanza, A., Mazzoni, A., Cicconetti, A., Testarelli, L., & Di Nardo, D. (2021). An update of the possible applications of magnetic resonance imaging (MRI) in dentistry: a literature review. *Journal of imaging*, *7*(5), 75. https://doi.org/10.3390/jimaging7050075.

Withers, P. J., Bouman, C., Carmignato, S., Cnudde, V., Grimaldi, D., Hagen, C. K., ... & Stock, S. R. (2021). X-ray computed tomography. *Nature Reviews Methods Primers*, *1*(1), 18. https://doi.org/10.1038/s43586-021-00015-4.

Nagesha, B. K., Kumar, S. A., Rajeswari, S., Barad, S., & Pathania, A. (2022). Inspection of additively manufactured aero-engine parts using computed radiography technique. *Journal of Materials Engineering and Performance*, *31*(8), 6322-6331. https://doi.org/10.1007/s11665-022-06890-6.

Posa, A., Tanzilli, A., Barbieri, P., Steri, L., Arbia, F., Mazza, G., ... & Iezzi, R. (2022). Digital subtraction angiography (DSA) technical and diagnostic aspects in the study of lower limb arteries. *Radiation*, *2*(4), 376-386. https://doi.org/10.3390/radiation2040028.

Elakkiya, S., & Thivya, K. S. (2022). Comprehensive review on lossy and lossless compression techniques. *Journal of The Institution of Engineers (India): Series B*, *103*, 1003-1012. https://doi.org/10.1007/s40031-021-00686-3.

Zerva, M. C., Christou, V., Giannakeas, N., Tzallas, A. T., & Kondi, L. P. (2023). An improved medical image compression method based on wavelet difference reduction. *IEEE Access*, *11*, 18026-18037. DOI: 10.1109/ACCESS.2023.3246948.

Ma, J. W., Czerniawski, T., & Leite, F. (2021). An application of metadata-based image retrieval system for facility management. *Advanced Engineering Informatics*, *50*, 101417. https://doi.org/10.1016/j.aei.2021.101417.

Song, J., Wang, B., Jiang, Q., & Hao, X. (2024). Exploring the Role of Fractal Geometry in Engineering Image Processing Based on Similarity and Symmetry: A Review. *Symmetry*, *16*(12), 1658. https://doi.org/10.3390/sym16121658.